

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070921\
 Data File : VD069649.D
 Acq On : 09 Jul 2021 16:32
 Operator : VA/SY
 Sample : M2965-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RO-ETGI-345

Quant Time: Jul 10 02:55:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 22 04:33:24 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	8814	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	15644	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	16030	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	5954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	6331	57.625	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.260%
35) Dibromofluoromethane	7.909	113	4762	45.031	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.060%
50) Toluene-d8	10.338	98	19067	48.069	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.140%
62) 4-Bromofluorobenzene	12.620	95	6017	46.141	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD070921\
Data File : VD069649.D
Acq On : 09 Jul 2021 16:32
Operator : VA/SY
Sample : M2965-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RO-ETGI-345

Quant Time: Jul 10 02:55:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D062121S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 22 04:33:24 2021
Response via : Initial Calibration

