


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD071019\
Data File : VD063055.D
Acq On : 10 Jul 2019 13:32
Operator : VA/AP
Sample : BFB
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D061919S.M
Title : SW846 8260
Last Update : Thu Jun 20 01:50:56 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	98080	PASS
75	95	30	60	46.8	209026	PASS
95	95	100	100	100.0	447018	PASS
96	95	5	9	5.9	26410	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	87.4	390656	PASS
175	174	5	9	6.9	27032	PASS
176	174	95	101	97.4	380352	PASS
177	176	5	9	6.4	24171	PASS