



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071122\
 Data File : VD073821.D
 Acq On : 11 Jul 2022 09:26
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071122S.M
 Title : SW846 8260
 Last Update : Tue Jul 12 07:04:44 2022



AutoFind: Scans 1627, 1628, 1629; Background Corrected with Scan 1616

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.4	15893	PASS
75	95	30	60	57.3	40624	PASS
95	95	100	100	100.0	70877	PASS
96	95	5	9	6.5	4611	PASS
173	174	0.00	2	1.6	765	PASS
174	95	50	100	66.1	46848	PASS
175	174	5	9	7.8	3677	PASS
176	174	95	101	96.6	45251	PASS
177	176	5	9	6.0	2713	PASS