

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071222\
 Data File : VD073835.D
 Acq On : 12 Jul 2022 13:29
 Operator : VA/SY
 Sample : N3645-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-08-070822

Quant Time: Jul 13 04:30:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071122S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 12 07:04:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	85679	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	200978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	183681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	74892	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	72544	63.213	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 126.420%	
35) Dibromofluoromethane	7.908	113	67699	50.919	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.840%	
50) Toluene-d8	10.332	98	253350	54.647	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	= 109.300%	
62) 4-Bromofluorobenzene	12.620	95	84801	43.169	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	= 86.340%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.991	85	1980	4.075	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071222\
Data File : VD073835.D
Acq On : 12 Jul 2022 13:29
Operator : VA/SY
Sample : N3645-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-08-070822

Quant Time: Jul 13 04:30:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071122S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 12 07:04:44 2022
Response via : Initial Calibration

