



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071222\
Data File : VD073830.D
Acq On : 12 Jul 2022 09:00
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071122S.M
Title : SW846 8260
Last Update : Tue Jul 12 07:04:44 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.7	18359	PASS
75	95	30	60	57.5	46440	PASS
95	95	100	100	100.0	80803	PASS
96	95	5	9	6.8	5522	PASS
173	174	0.00	2	1.6	803	PASS
174	95	50	100	63.2	51059	PASS
175	174	5	9	7.5	3814	PASS
176	174	95	101	96.7	49387	PASS
177	176	5	9	6.4	3174	PASS