

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071223\
 Data File : VD076726.D
 Acq On : 12 Jul 2023 11:45
 Operator : KP/SY
 Sample : 03581-01
 Misc : 5.03G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-1

Quant Time: Jul 12 12:24:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 19 04:51:45 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	101082	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	246696	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	263189	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	122540	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	68743	67.877	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.760%
35) Dibromofluoromethane	7.810	113	82560	56.334	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.660%
50) Toluene-d8	10.269	98	321170	57.785	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	115.580%
62) 4-Bromofluorobenzene	12.575	95	113451	62.895	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	125.800%
Target Compounds						
16) Acetone	4.016	43	2206	9.154	ug/l	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071223\
Data File : VD076726.D
Acq On : 12 Jul 2023 11:45
Operator : KP/SY
Sample : 03581-01
Misc : 5.03G/5.0ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-1

Quant Time: Jul 12 12:24:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D061623S.M
Quant Title : SW846 8260
QLast Update : Mon Jun 19 04:51:45 2023
Response via : Initial Calibration

