

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071321\
 Data File : VD069679.D
 Acq On : 13 Jul 2021 19:27
 Operator : VA/SY
 Sample : M2998-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-231

Quant Time: Jul 14 06:08:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	590036	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1200194	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	1343914	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	618273	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	356920	48.446	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.900%
35) Dibromofluoromethane	7.914	113	365869	45.439	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.880%
50) Toluene-d8	10.338	98	1595236	50.996	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.000%
62) 4-Bromofluorobenzene	12.626	95	643388	61.670	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	123.340%
Target Compounds						
16) Acetone	4.162	43	9921	18.909	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071321\
Data File : VD069679.D
Acq On : 13 Jul 2021 19:27
Operator : VA/SY
Sample : M2998-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-231

Quant Time: Jul 14 06:08:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

