



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071321\
Data File : VD069661.D
Acq On : 13 Jul 2021 09:32
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Title : SW846 8260
Last Update : Wed Jul 14 06:00:31 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.0	72928	PASS
75	95	30	60	49.1	210475	PASS
95	95	100	100	100.0	428331	PASS
96	95	5	9	6.4	27275	PASS
173	174	0.00	2	1.2	3683	PASS
174	95	50	100	69.9	299221	PASS
175	174	5	9	7.2	21656	PASS
176	174	95	101	96.1	287637	PASS
177	176	5	9	6.7	19184	PASS