

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
 Data File : VD069688.D
 Acq On : 14 Jul 2021 15:17
 Operator : VA/SY
 Sample : M3001-03
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 N9994

Quant Time: Jul 15 10:02:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	537592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1094672	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1192931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	539018	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	325031	48.422	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.840%
35) Dibromofluoromethane	7.914	113	347777	47.356	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.720%
50) Toluene-d8	10.338	98	1435357	50.308	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.620%
62) 4-Bromofluorobenzene	12.626	95	567512	59.641	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	119.280%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.185	59	16816	18.750	ug/l #	79
16) Acetone	4.156	43	29754	32.236	ug/l	94
20) Methylene Chloride	4.962	84	76961	3.611	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
Data File : VD069688.D
Acq On : 14 Jul 2021 15:17
Operator : VA/SY
Sample : M3001-03
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
N9994

Quant Time: Jul 15 10:02:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

