

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
 Data File : VD069695.D
 Acq On : 14 Jul 2021 18:37
 Operator : VA/SY
 Sample : M3023-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-071321

Quant Time: Jul 15 03:32:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	64462	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	123998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	154314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	71196	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	63358	78.716	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.440%
35) Dibromofluoromethane	7.914	113	44755	53.800	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.600%
50) Toluene-d8	10.337	98	162528	50.290	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.580%
62) 4-Bromofluorobenzene	12.626	95	70782	65.669	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	131.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
Data File : VD069695.D
Acq On : 14 Jul 2021 18:37
Operator : VA/SY
Sample : M3023-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-071321

Quant Time: Jul 15 03:32:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

