

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
 Data File : VD069698.D
 Acq On : 14 Jul 2021 20:03
 Operator : VA/SY
 Sample : M3026-01
 Misc : 11.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ABN-JG-1

Quant Time: Jul 15 03:33:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	452709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	963016	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1201940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	546738	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	322668	57.083	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.160%
35) Dibromofluoromethane	7.908	113	76891	11.901	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	23.800%#
50) Toluene-d8	10.338	98	1370519	54.603	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.200%
62) 4-Bromofluorobenzene	12.626	95	599217	71.582	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	143.160%
Target Compounds						
					Qvalue	
16) Acetone	4.161	43	44665	47.443	ug/l	99
17) Carbon Disulfide	4.403	76	10849	1.053	ug/l	100
25) 2-Butanone	7.214	43	8451	6.326	ug/l	97
39) Methylcyclohexane	9.355	83	29671	4.361	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
Data File : VD069698.D
Acq On : 14 Jul 2021 20:03
Operator : VA/SY
Sample : M3026-01
Misc : 11.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ABN-JG-1

Quant Time: Jul 15 03:33:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

