

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
 Data File : VD069700.D
 Acq On : 14 Jul 2021 21:00
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 03:33:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	81	0.00
2 T	Dichlorodifluoromethane	0.498	0.565	-13.5	81	0.00
3 P	Chloromethane	0.705	0.828	-17.4	89	0.00
4 C	Vinyl Chloride	0.761	0.846	-11.2#	84	0.00
5 T	Bromomethane	0.500	0.528	-5.6	89	0.00
6 T	Chloroethane	0.487	0.552	-13.3	88	0.00
7 T	Trichlorofluoromethane	0.959	1.002	-4.5	81	0.00
8 T	Diethyl Ether	0.280	0.326	-16.4	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.595	0.632	-6.2	83	0.00
10 T	Methyl Iodide	0.554	0.624	-12.6	79	0.00
11 T	Tert butyl alcohol	0.083	0.051	38.6#	92	0.01
12 CM	1,1-Dichloroethene	0.545	0.607	-11.4#	82	0.00
13 T	Acrolein	0.043	0.041	4.7	83	0.00
14 T	Allyl chloride	0.843	0.936	-11.0	84	0.00
15 T	Acrylonitrile	0.127	0.150	-18.1	92	0.00
16 T	Acetone	0.114	0.105	7.9	78	0.00
17 T	Carbon Disulfide	1.857	2.176	-17.2	83	0.00
18 T	Methyl Acetate	0.352	0.328	6.8	93	0.00
19 T	Methyl tert-butyl Ether	1.154	1.426	-23.6	90	0.00
20 T	Methylene Chloride	1.096	0.997	9.0	98	0.00
21 T	trans-1,2-Dichloroethene	0.651	0.729	-12.0	86	0.00
22 T	Diisopropyl ether	1.746	2.251	-28.9#	92	0.00
23 T	Vinyl Acetate	0.828	1.114	-34.5#	90	0.00
24 P	1,1-Dichloroethane	1.219	1.409	-15.6	90	0.00
25 T	2-Butanone	0.148	0.169	-14.2	91	0.00
26 T	2,2-Dichloropropane	0.972	0.948	2.5	78	0.00
27 T	cis-1,2-Dichloroethene	0.702	0.830	-18.2	91	0.00
28 T	Bromochloromethane	0.470	0.568	-20.9	94	0.00
29 T	Tetrahydrofuran	0.091	0.107	-17.6	87	0.00
30 C	Chloroform	1.212	1.311	-8.2#	86	0.00
31 T	Cyclohexane	1.037	1.066	-2.8	76	0.00
32 T	1,1,1-Trichloroethane	1.028	1.040	-1.2	80	0.00
33 S	1,2-Dichloroethane-d4	0.624	0.717	-14.9	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.335	0.343	-2.4	92	0.00
36 T	1,1-Dichloropropene	0.516	0.490	5.0	80	0.00
37 T	Ethyl Acetate	0.195	0.207	-6.2	92	0.00
38 T	Carbon Tetrachloride	0.472	0.436	7.6	79	0.00
39 T	Methylcyclohexane	0.559	0.549	1.8	76	0.00
40 TM	Benzene	1.488	1.607	-8.0	92	0.00
41 T	Methacrylonitrile	0.101	0.106	-5.0	92	0.00
42 TM	1,2-Dichloroethane	0.404	0.424	-5.0	93	0.00
43 T	Isopropyl Acetate	0.357	0.380	-6.4	93	0.00
44 TM	Trichloroethene	0.361	0.352	2.5	85	0.00
45 C	1,2-Dichloropropane	0.391	0.395	-1.0#	88	0.00
46 T	Dibromomethane	0.191	0.203	-6.3	93	0.00
47 T	Bromodichloromethane	0.506	0.516	-2.0	90	0.00
48 T	Methyl methacrylate	0.169	0.182	-7.7	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
 Data File : VD069700.D
 Acq On : 14 Jul 2021 21:00
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 03:33:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	97	0.00
50 S	Toluene-d8	1.303	1.368	-5.0	91	0.00
51 T	4-Methyl-2-Pentanone	0.185	0.208	-12.4	95	0.00
52 CM	Toluene	0.895	0.992	-10.8#	91	0.00
53 T	t-1,3-Dichloropropene	0.425	0.433	-1.9	89	0.00
54 T	cis-1,3-Dichloropropene	0.522	0.521	0.2	90	0.00
55 T	1,1,2-Trichloroethane	0.267	0.267	0.0	89	0.00
56 T	Ethyl methacrylate	0.305	0.348	-14.1	92	0.00
57 T	1,3-Dichloropropane	0.471	0.494	-4.9	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.136	0.176	-29.4#	110	0.00
59 T	2-Hexanone	0.123	0.131	-6.5	88	0.00
60 T	Dibromochloromethane	0.311	0.338	-8.7	97	0.00
61 T	1,2-Dibromoethane	0.248	0.267	-7.7	96	0.00
62 S	4-Bromofluorobenzene	0.435	0.474	-9.0	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.305	0.302	1.0	91	0.00
65 PM	Chlorobenzene	0.999	1.030	-3.1	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.366	-2.5	94	0.00
67 C	Ethyl Benzene	1.769	1.876	-6.0#	91	0.00
68 T	m/p-Xylenes	0.669	0.724	-8.2	92	0.00
69 T	o-Xylene	0.609	0.668	-9.7	93	0.00
70 T	Styrene	1.063	1.190	-11.9	94	0.00
71 P	Bromoform	0.173	0.185	-6.9	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.560	3.589	-0.8	91	0.00
74 T	N-amyl acetate	0.774	0.782	-1.0	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.709	0.706	0.4	99	0.00
76 T	1,2,3-Trichloropropane	0.486	0.485	0.2	97	0.00
77 T	Bromobenzene	0.790	0.797	-0.9	96	0.00
78 T	n-propylbenzene	4.514	4.630	-2.6	91	0.00
79 T	2-Chlorotoluene	2.654	2.685	-1.2	93	0.00
80 T	1,3,5-Trimethylbenzene	3.020	3.137	-3.9	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.181	0.174	3.9	96	0.00
82 T	4-Chlorotoluene	2.785	2.812	-1.0	94	0.00
83 T	tert-Butylbenzene	2.462	2.483	-0.9	92	0.00
84 T	1,2,4-Trimethylbenzene	2.996	3.139	-4.8	94	0.00
85 T	sec-Butylbenzene	3.869	3.871	-0.1	89	0.00
86 T	p-Isopropyltoluene	3.151	3.218	-2.1	91	0.00
87 T	1,3-Dichlorobenzene	1.640	1.655	-0.9	96	0.00
88 T	1,4-Dichlorobenzene	1.638	1.620	1.1	95	0.00
89 T	n-Butylbenzene	3.122	3.099	0.7	89	0.00
90 T	Hexachloroethane	0.641	0.627	2.2	93	0.00
91 T	1,2-Dichlorobenzene	1.414	1.443	-2.1	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.106	0.101	4.7	95	0.00
93 T	1,2,4-Trichlorobenzene	0.786	0.812	-3.3	97	0.00
94 T	Hexachlorobutadiene	0.474	0.453	4.4	91	0.00
95 T	Naphthalene	1.426	1.513	-6.1	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071421\
Data File : VD069700.D
Acq On : 14 Jul 2021 21:00
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jul 15 03:33:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.675	0.705	-4.4	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6