


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD071519\
 Data File : VD063074.D
 Acq On : 15 Jul 2019 9:24
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D071519S.M
 Title : SW846 8260
 Last Update : Tue Jul 16 08:11:56 2019



AutoFind: Scans 1230, 1231, 1232; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	81096	PASS
75	95	30	60	47.8	173472	PASS
95	95	100	100	100.0	362986	PASS
96	95	5	9	6.0	21813	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.2	312941	PASS
175	174	5	9	6.8	21321	PASS
176	174	95	101	95.2	298026	PASS
177	176	5	9	6.0	17985	PASS