

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071521\
 Data File : VD069705.D
 Acq On : 15 Jul 2021 14:05
 Operator : VA/SY
 Sample : M3014-01
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-02-071221

Quant Time: Jul 15 15:52:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	544497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1103748	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	1206818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	506056	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	336512	49.496	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.000%
35) Dibromofluoromethane	7.914	113	356119	48.093	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.180%
50) Toluene-d8	10.337	98	1443482	50.177	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.360%
62) 4-Bromofluorobenzene	12.626	95	549808	57.305	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.620%
Target Compounds						
16) Acetone	4.155	43	36702	36.445	ug/l #	85
20) Methylene Chloride	4.961	84	89466	4.866	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071521\
Data File : VD069705.D
Acq On : 15 Jul 2021 14:05
Operator : VA/SY
Sample : M3014-01
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-02-071221

Quant Time: Jul 15 15:52:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

