

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071521\
 Data File : VD069708.D
 Acq On : 15 Jul 2021 15:38
 Operator : VA/SY
 Sample : M3002-01
 Misc : 5.56G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-F

Quant Time: Jul 16 02:02:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	545824	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1100836	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1161642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	454377	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	325682	47.787	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.580%
35) Dibromofluoromethane	7.914	113	351115	47.542	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.080%
50) Toluene-d8	10.338	98	1429096	49.808	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.620%
62) 4-Bromofluorobenzene	12.626	95	506512	52.932	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	105.860%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	51527	45.937	ug/l	96
17) Carbon Disulfide	4.403	76	15983	1.190	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071521\
Data File : VD069708.D
Acq On : 15 Jul 2021 15:38
Operator : VA/SY
Sample : M3002-01
Misc : 5.56G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-F

Quant Time: Jul 16 02:02:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

