

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071521\
 Data File : VD069702.D
 Acq On : 15 Jul 2021 09:58
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 15:50:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.498	0.569	-14.3	96	0.00
3 P	Chloromethane	0.705	0.748	-6.1	95	0.00
4 C	Vinyl Chloride	0.761	0.832	-9.3#	97	0.00
5 T	Bromomethane	0.500	0.508	-1.6	101	0.00
6 T	Chloroethane	0.487	0.515	-5.7	97	0.00
7 T	Trichlorofluoromethane	0.959	1.026	-7.0	98	0.00
8 T	Diethyl Ether	0.280	0.280	0.0	92	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.595	0.650	-9.2	101	0.00
10 T	Methyl Iodide	0.554	0.602	-8.7	90	0.00
11 T	Tert butyl alcohol	0.083	0.027	67.5#	56	0.00
12 CM	1,1-Dichloroethene	0.545	0.594	-9.0#	95	0.00
13 T	Acrolein	0.043	0.056	-30.2#	133	0.00
14 T	Allyl chloride	0.843	0.855	-1.4	90	0.00
15 T	Acrylonitrile	0.127	0.134	-5.5	97	0.00
16 T	Acetone	0.114	0.128	-12.3	112	0.00
17 T	Carbon Disulfide	1.857	2.103	-13.2	95	0.00
18 T	Methyl Acetate	0.352	0.293	16.8	98	0.00
19 T	Methyl tert-butyl Ether	1.154	1.250	-8.3	93	0.00
20 T	Methylene Chloride	1.096	0.784	28.5#	91	0.00
21 T	trans-1,2-Dichloroethene	0.651	0.687	-5.5	96	0.00
22 T	Diisopropyl ether	1.746	1.931	-10.6	93	0.00
23 T	Vinyl Acetate	0.828	0.976	-17.9	93	0.00
24 P	1,1-Dichloroethane	1.219	1.261	-3.4	95	0.00
25 T	2-Butanone	0.148	0.157	-6.1	100	0.00
26 T	2,2-Dichloropropane	0.972	1.009	-3.8	98	0.00
27 T	cis-1,2-Dichloroethene	0.702	0.724	-3.1	94	0.00
28 T	Bromochloromethane	0.470	0.546	-16.2	107	0.00
29 T	Tetrahydrofuran	0.091	0.099	-8.8	95	0.00
30 C	Chloroform	1.212	1.246	-2.8#	96	0.00
31 T	Cyclohexane	1.037	1.150	-10.9	97	0.00
32 T	1,1,1-Trichloroethane	1.028	1.061	-3.2	97	0.00
33 S	1,2-Dichloroethane-d4	0.624	0.607	2.7	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.335	0.360	-7.5	97	0.00
36 T	1,1-Dichloropropene	0.516	0.595	-15.3	98	0.00
37 T	Ethyl Acetate	0.195	0.215	-10.3	96	0.00
38 T	Carbon Tetrachloride	0.472	0.546	-15.7	100	0.00
39 T	Methylcyclohexane	0.559	0.692	-23.8	97	0.00
40 TM	Benzene	1.488	1.653	-11.1	95	0.00
41 T	Methacrylonitrile	0.101	0.112	-10.9	98	0.00
42 TM	1,2-Dichloroethane	0.404	0.439	-8.7	97	0.00
43 T	Isopropyl Acetate	0.357	0.388	-8.7	96	0.00
44 TM	Trichloroethene	0.361	0.395	-9.4	96	0.00
45 C	1,2-Dichloropropane	0.391	0.434	-11.0#	97	0.00
46 T	Dibromomethane	0.191	0.204	-6.8	94	0.00
47 T	Bromodichloromethane	0.506	0.553	-9.3	98	0.00
48 T	Methyl methacrylate	0.169	0.192	-13.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071521\
 Data File : VD069702.D
 Acq On : 15 Jul 2021 09:58
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 15 15:50:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	94	0.00
50 S	Toluene-d8	1.303	1.421	-9.1	96	0.00
51 T	4-Methyl-2-Pentanone	0.185	0.214	-15.7	99	0.00
52 CM	Toluene	0.895	1.046	-16.9#	97	0.00
53 T	t-1,3-Dichloropropene	0.425	0.471	-10.8	97	0.00
54 T	cis-1,3-Dichloropropene	0.522	0.554	-6.1	96	0.00
55 T	1,1,2-Trichloroethane	0.267	0.295	-10.5	99	0.00
56 T	Ethyl methacrylate	0.305	0.359	-17.7	96	0.00
57 T	1,3-Dichloropropane	0.471	0.522	-10.8	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.136	0.168	-23.5	107	0.00
59 T	2-Hexanone	0.123	0.151	-22.8	103	0.00
60 T	Dibromochloromethane	0.311	0.347	-11.6	100	0.00
61 T	1,2-Dibromoethane	0.248	0.278	-12.1	101	0.00
62 S	4-Bromofluorobenzene	0.435	0.477	-9.7	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.305	0.331	-8.5	97	0.00
65 PM	Chlorobenzene	0.999	1.093	-9.4	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.397	-11.2	100	0.00
67 C	Ethyl Benzene	1.769	2.044	-15.5#	96	0.00
68 T	m/p-Xylenes	0.669	0.805	-20.3	99	0.00
69 T	o-Xylene	0.609	0.717	-17.7	97	0.00
70 T	Styrene	1.063	1.281	-20.5	98	0.00
71 P	Bromoform	0.173	0.197	-13.9	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.560	3.981	-11.8	98	0.00
74 T	N-amyl acetate	0.774	0.830	-7.2	97	0.00
75 P	1,1,2,2-Tetrachloroethane	0.709	0.755	-6.5	103	0.00
76 T	1,2,3-Trichloropropane	0.486	0.497	-2.3	96	0.00
77 T	Bromobenzene	0.790	0.851	-7.7	99	0.00
78 T	n-propylbenzene	4.514	5.196	-15.1	99	0.00
79 T	2-Chlorotoluene	2.654	2.888	-8.8	97	0.00
80 T	1,3,5-Trimethylbenzene	3.020	3.428	-13.5	99	0.00
81 T	trans-1,4-Dichloro-2-butene	0.181	0.201	-11.0	108	0.00
82 T	4-Chlorotoluene	2.785	3.038	-9.1	99	0.00
83 T	tert-Butylbenzene	2.462	2.794	-13.5	100	0.00
84 T	1,2,4-Trimethylbenzene	2.996	3.374	-12.6	98	0.00
85 T	sec-Butylbenzene	3.869	4.412	-14.0	99	0.00
86 T	p-Isopropyltoluene	3.151	3.650	-15.8	101	0.00
87 T	1,3-Dichlorobenzene	1.640	1.797	-9.6	101	0.00
88 T	1,4-Dichlorobenzene	1.638	1.727	-5.4	99	0.00
89 T	n-Butylbenzene	3.122	3.567	-14.3	100	0.00
90 T	Hexachloroethane	0.641	0.701	-9.4	101	0.00
91 T	1,2-Dichlorobenzene	1.414	1.535	-8.6	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.106	0.110	-3.8	101	0.00
93 T	1,2,4-Trichlorobenzene	0.786	0.849	-8.0	98	0.00
94 T	Hexachlorobutadiene	0.474	0.521	-9.9	102	0.00
95 T	Naphthalene	1.426	1.582	-10.9	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071521\
Data File : VD069702.D
Acq On : 15 Jul 2021 09:58
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Jul 15 15:50:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.675	0.744	-10.2	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6