

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD071619\
 Data File : VD063108.D
 Acq On : 16 Jul 2019 15:59
 Operator : VA/AP
 Sample : K3776-04
 Misc : 4.54g/5ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-SF-01-006-B

Quant Time: Jul 17 08:13:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D071519S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 16 08:11:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	507208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	704578	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.36	117	445477	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	127907	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	346989	54.92	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.84%	
35) Dibromofluoromethane	6.91	113	66661	9.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	19.00%	
50) Toluene-d8	10.40	98	587613	38.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	77.96%	
62) 4-Bromofluorobenzene	13.55	95	205365	29.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	58.90%	
Target Compounds						
16) Acetone	3.21	43	57407	37.618	ug/l	95
80) 1,3,5-Trimethylbenzene	13.93	105	32448	3.769	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD071619\
Data File : VD063108.D
Acq On : 16 Jul 2019 15:59
Operator : VA/AP
Sample : K3776-04
Misc : 4.54g/5ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-SF-01-006-B

Quant Time: Jul 17 08:13:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D071519S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 16 08:11:56 2019
Response via : Initial Calibration

