

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD071620\
 Data File : VD065965.D
 Acq On : 16 Jul 2020 18:56
 Operator : VA/SY
 Sample : L3308-05
 Misc : 5.03G/5.00ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 5

Quant Time: Jul 17 04:41:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D070720S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 08 09:45:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	353438	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	606977	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	622896	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	296716	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	166228	53.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.14%	
35) Dibromofluoromethane	7.91	113	189463	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
50) Toluene-d8	10.34	98	738324	54.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.08%	
62) 4-Bromofluorobenzene	12.64	95	271695	57.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.82%	
Target Compounds						
8) Diethyl Ether	3.71	74	5027	2.989	ug/l	88
16) Acetone	4.16	43	34825	47.441	ug/l	99
20) Methylene Chloride	4.96	84	118765	25.007	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD071620\
Data File : VD065965.D
Acq On : 16 Jul 2020 18:56
Operator : VA/SY
Sample : L3308-05
Misc : 5.03G/5.00ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
5

Quant Time: Jul 17 04:41:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D070720S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 08 09:45:47 2020
Response via : Initial Calibration

