

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071621\
 Data File : VD069735.D
 Acq On : 16 Jul 2021 19:56
 Operator : VA/SY
 Sample : M3022-02
 Misc : 5.51G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ETGI-363

Quant Time: Jul 17 00:56:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	497587	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	968206	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1186529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	544928	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	336987	54.239	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	108.480%	
35) Dibromofluoromethane	7.909	113	318518	49.037	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	98.080%	
50) Toluene-d8	10.338	98	1378035	54.608	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	109.220%	
62) 4-Bromofluorobenzene	12.626	95	556656	66.141	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	132.280%	
Target Compounds						
17) Carbon Disulfide	4.409	76	12020	1.058	ug/l	Qvalue # 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071621\
Data File : VD069735.D
Acq On : 16 Jul 2021 19:56
Operator : VA/SY
Sample : M3022-02
Misc : 5.51G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ETGI-363

Quant Time: Jul 17 00:56:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

