


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071621\
 Data File : VD069716.D
 Acq On : 16 Jul 2021 09:40
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Title : SW846 8260
 Last Update : Wed Jul 14 06:00:31 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1806

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.4	83341	PASS
75	95	30	60	48.4	232320	PASS
95	95	100	100	100.0	480235	PASS
96	95	5	9	6.7	32032	PASS
173	174	0.00	2	1.2	4088	PASS
174	95	50	100	69.8	335168	PASS
175	174	5	9	7.5	25091	PASS
176	174	95	101	98.3	329323	PASS
177	176	5	9	6.2	20469	PASS