

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071723\
 Data File : VD076755.D
 Acq On : 17 Jul 2023 15:29
 Operator : KP/SY
 Sample : 03645-05
 Misc : 4.29G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-09-(2.0-4.0)

Quant Time: Jul 18 01:23:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	163319	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	370057	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	357989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	152519	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	103720	63.761	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 127.520%	
35) Dibromofluoromethane	7.805	113	124112	53.371	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.740%	
50) Toluene-d8	10.269	98	460958	55.410	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 110.820%	
62) 4-Bromofluorobenzene	12.575	95	137641	51.144	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 102.280%	
Target Compounds						
					Qvalue	
31) Cyclohexane	7.863	56	3686	1.540	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071723\
Data File : VD076755.D
Acq On : 17 Jul 2023 15:29
Operator : KP/SY
Sample : 03645-05
Misc : 4.29G/5.0ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-09-(2.0-4.0)

Quant Time: Jul 18 01:23:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

