

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071723\
 Data File : VD076758.D
 Acq On : 17 Jul 2023 16:54
 Operator : KP/SY
 Sample : 03638-01
 Misc : 5.12G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP1

Quant Time: Jul 18 01:24:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	149314	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	352807	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	346969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	156240	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	95866	64.460	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 128.920%	
35) Dibromofluoromethane	7.804	113	119328	53.822	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.640%	
50) Toluene-d8	10.269	98	441099	55.615	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 111.240%	
62) 4-Bromofluorobenzene	12.575	95	139996	54.563	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 109.120%	
Target Compounds						
						Qvalue
17) Carbon Disulfide	4.263	76	4384	1.272	ug/l	# 87
31) Cyclohexane	7.875	56	3401	1.554	ug/l	# 48

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071723\
Data File : VD076758.D
Acq On : 17 Jul 2023 16:54
Operator : KP/SY
Sample : 03638-01
Misc : 5.12G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP1

Quant Time: Jul 18 01:24:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

