

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071822\
 Data File : VD073854.D
 Acq On : 18 Jul 2022 17:39
 Operator : VA/SY
 Sample : VIBLK
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Jul 19 02:00:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 01:56:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	123525	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	254552	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	238931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	106732	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	81274	55.100	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.200%
35) Dibromofluoromethane	7.909	113	88987	49.552	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.100%
50) Toluene-d8	10.332	98	316464	48.543	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.080%
62) 4-Bromofluorobenzene	12.620	95	106993	47.734	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	95.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071822\
Data File : VD073854.D
Acq On : 18 Jul 2022 17:39
Operator : VA/SY
Sample : VIBLK
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Jul 19 02:00:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 01:56:16 2022
Response via : Initial Calibration

