



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071822\
Data File : VD073847.D
Acq On : 18 Jul 2022 09:04
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Title : SW846 8260
Last Update : Tue Jul 19 01:56:16 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.9	14499	PASS
75	95	30	60	56.0	38885	PASS
95	95	100	100	100.0	69392	PASS
96	95	5	9	6.0	4132	PASS
173	174	0.00	2	0.9	465	PASS
174	95	50	100	76.6	53152	PASS
175	174	5	9	7.6	4063	PASS
176	174	95	101	95.6	50808	PASS
177	176	5	9	6.9	3520	PASS