

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071823\
 Data File : VD076773.D
 Acq On : 18 Jul 2023 12:56
 Operator : KP/SY
 Sample : 03654-01
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 72-12016

Quant Time: Jul 19 06:42:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	149773	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	358518	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	346278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	144567	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	98026	65.711	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.420%
35) Dibromofluoromethane	7.799	113	117944	52.351	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.700%
50) Toluene-d8	10.269	98	450448	55.889	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	111.780%
62) 4-Bromofluorobenzene	12.575	95	129930	49.833	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.660%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071823\
Data File : VD076773.D
Acq On : 18 Jul 2023 12:56
Operator : KP/SY
Sample : 03654-01
Misc : 5.05G/5.0ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
72-12016

Quant Time: Jul 19 06:42:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD076773.D
Acq: 18 Jul 2023 12:56




Instrument :
MSVOA_D
ClientSampleId :
72-12016


Tgt Ion:168 Resp: 149773
Ion Ratio Lower Upper
168 100
99 55.4 36.7 55.1#



#33
1,2-Dichloroethane-d4
Concen: 65.711 ug/l
RT: 8.228 min Scan# 1129
Delta R.T. -0.000 min
Lab File: VD076773.D
Acq: 18 Jul 2023 12:56

Tgt Ion: 65 Resp: 98026
Ion Ratio Lower Upper
65 100
67 55.4 0.0 115.0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1

Delta R.T. -0.000 min

Lab File: VD076773.D

Acq: 18 Jul 2023 12:56

Instrument :

MSVOA_D

ClientSampleId :

72-12016





Tgt Ion:114 Resp: 358518

Ion Ratio Lower Upper

114 100

63 17.7 0.0 33.4

88 13.3 0.0 26.2



#35

Dibromofluoromethane

Concen: 52.351 ug/l

RT: 7.799 min Scan# 1056

Delta R.T. -0.006 min

Lab File: VD076773.D

Acq: 18 Jul 2023 12:56




Tgt Ion:113 Resp: 117944


Ion Ratio Lower Upper

113 100

111 103.9 84.7 127.1

192 17.5 16.8 25.2









#50
Toluene-d8
Concen: 55.889 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD076773.D
Acq: 18 Jul 2023 12:56




Instrument :
MSVOA_D
ClientSampleId :
72-12016


Tgt Ion: 98 Resp: 450448
Ion Ratio Lower Upper
98 100
100 65.8 50.3 75.5



#62
4-Bromofluorobenzene
Concen: 49.833 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD076773.D
Acq: 18 Jul 2023 12:56

Tgt Ion: 95 Resp: 129930
Ion Ratio Lower Upper
95 100
174 85.5 0.0 201.8
176 81.8 0.0 191.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD076773.D
Acq: 18 Jul 2023 12:56

Instrument :
MSVOA_D
ClientSampleId :
72-12016

Tgt Ion:117 Resp: 346278
Ion Ratio Lower Upper
117 100
82 49.2 38.3 57.5
119 31.6 25.0 37.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.006 min
Lab File: VD076773.D
Acq: 18 Jul 2023 12:56

Tgt Ion:152 Resp: 144567
Ion Ratio Lower Upper
152 100
115 59.2 27.1 81.3
150 159.8 0.0 353.6

