


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD071919\
 Data File : VD063145.D
 Acq On : 19 Jul 2019 9:03
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D071519S.M
 Title : SW846 8260
 Last Update : Tue Jul 16 08:11:56 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1226

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.4	89309	PASS
75	95	30	60	50.0	190832	PASS
95	95	100	100	100.0	381760	PASS
96	95	5	9	6.4	24550	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.4	337602	PASS
175	174	5	9	7.0	23539	PASS
176	174	95	101	95.2	321360	PASS
177	176	5	9	6.2	20034	PASS