

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071921\
 Data File : VD069740.D
 Acq On : 19 Jul 2021 09:31
 Operator : VA/SY
 Sample : VD0719SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0719SBL01

Quant Time: Jul 20 02:21:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	591866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1131045	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1317812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	618469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	346968	46.950	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	93.900%		
35) Dibromofluoromethane	7.914	113	359305	47.352	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	94.700%		
50) Toluene-d8	10.338	98	1494804	50.707	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	101.420%		
62) 4-Bromofluorobenzene	12.626	95	634360	64.522	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	129.040%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.991	85	3752	1.222	ug/l	91
16) Acetone	4.162	43	8152	17.865	ug/l	90
17) Carbon Disulfide	4.409	76	21218	1.359	ug/l #	94
20) Methylene Chloride	4.956	84	59628	1.111	ug/l	97
39) Methylcyclohexane	9.350	83	5873	2.363	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071921\
Data File : VD069740.D
Acq On : 19 Jul 2021 09:31
Operator : VA/SY
Sample : VD0719SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0719SBL01

Quant Time: Jul 20 02:21:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

