

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071921\
 Data File : VD069745.D
 Acq On : 19 Jul 2021 12:03
 Operator : VA/SY
 Sample : M3038-01
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-5.5-6.0

Quant Time: Jul 20 02:10:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	3715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	6254	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	6799	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	2541	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	3307	71.292	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.580%
35) Dibromofluoromethane	7.920	113	2188	52.149	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.300%
50) Toluene-d8	10.337	98	8890	54.539	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.080%
62) 4-Bromofluorobenzene	12.626	95	2928	53.860	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071921\
Data File : VD069745.D
Acq On : 19 Jul 2021 12:03
Operator : VA/SY
Sample : M3038-01
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-5.5-6.0

Quant Time: Jul 20 02:10:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

