

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071921\
 Data File : VD069748.D
 Acq On : 19 Jul 2021 13:29
 Operator : VA/SY
 Sample : M3065-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-04-071521

Quant Time: Jul 20 02:11:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	1404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	2107	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1542	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.57
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	505	28.807	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	57.620%
35) Dibromofluoromethane	7.909	113	741	52.421	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.840%
50) Toluene-d8	10.338	98	2353	42.847	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	85.700%
62) 4-Bromofluorobenzene	12.620	95	178	9.719	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	19.440%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071921\
Data File : VD069748.D
Acq On : 19 Jul 2021 13:29
Operator : VA/SY
Sample : M3065-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-04-071521

Quant Time: Jul 20 02:11:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

