

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071922\
 Data File : VD073861.D
 Acq On : 19 Jul 2022 15:10
 Operator : VA/SY
 Sample : N3688-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 72-12016

Quant Time: Jul 19 15:44:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 01:56:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	117285	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	233286	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	213521	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	94116	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	64583	46.114	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.220%
35) Dibromofluoromethane	7.902	113	77783	47.262	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.520%
50) Toluene-d8	10.332	98	288268	48.249	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.500%
62) 4-Bromofluorobenzene	12.620	95	92294	44.930	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	89.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071922\
Data File : VD073861.D
Acq On : 19 Jul 2022 15:10
Operator : VA/SY
Sample : N3688-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
72-12016

Quant Time: Jul 19 15:44:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 01:56:16 2022
Response via : Initial Calibration

