

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071922\
 Data File : VD073871.D
 Acq On : 19 Jul 2022 20:15
 Operator : VA/SY
 Sample : N3756-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1

Quant Time: Jul 20 01:23:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 01:56:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	22122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	46217	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	46588	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	23126	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	21790	82.487	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 164.980%#	
35) Dibromofluoromethane	7.908	113	17942	55.028	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.060%	
50) Toluene-d8	10.332	98	55596	46.970	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.940%	
62) 4-Bromofluorobenzene	12.620	95	21499	52.828	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 105.660%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.997	85	646	1.167	ug/l	99
16) Acetone	4.161	43	3976	91.727	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071922\
Data File : VD073871.D
Acq On : 19 Jul 2022 20:15
Operator : VA/SY
Sample : N3756-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1

Quant Time: Jul 20 01:23:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 01:56:16 2022
Response via : Initial Calibration

