

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071922\
 Data File : VD073869.D
 Acq On : 19 Jul 2022 19:13
 Operator : VA/SY
 Sample : N3730-07
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-04-6.5-7.0

Quant Time: Jul 20 01:23:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 19 01:56:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	49777	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	107322	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	106957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	53192	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	46909	78.919	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	157.840%
35) Dibromofluoromethane	7.903	113	40482	53.467	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.940%
50) Toluene-d8	10.332	98	131410	47.810	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.620%
62) 4-Bromofluorobenzene	12.620	95	52711	55.778	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.560%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071922\
Data File : VD073869.D
Acq On : 19 Jul 2022 19:13
Operator : VA/SY
Sample : N3730-07
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-04-6.5-7.0

Quant Time: Jul 20 01:23:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071822S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 19 01:56:16 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.967 min Scan# 1028
Delta R.T. 0.000 min
Lab File: VD073869.D
Acq: 19 Jul 2022 19:13




Instrument :
MSVOA_D
ClientSampleId :
SB-04-6.5-7.0


Tgt Ion:168 Resp: 49777
Ion Ratio Lower Upper
168 100
99 64.0 49.8 74.6



#33
1,2-Dichloroethane-d4
Concen: 78.919 ug/l
RT: 8.320 min Scan# 1088
Delta R.T. -0.006 min
Lab File: VD073869.D
Acq: 19 Jul 2022 19:13

Tgt Ion: 65 Resp: 46909
Ion Ratio Lower Upper
65 100
67 50.8 0.0 101.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.855 min Scan# 1179

Delta R.T. -0.000 min

Lab File: VD073869.D

Acq: 19 Jul 2022 19:13

Instrument :

MSVOA_D

ClientSampleId :

SB-04-6.5-7.0





Tgt Ion:114 Resp: 107322

Ion Ratio Lower Upper

114 100

63 20.4 0.0 46.6

88 17.4 0.0 33.2



#35

Dibromofluoromethane

Concen: 53.467 ug/l

RT: 7.903 min Scan# 1017

Delta R.T. -0.006 min

Lab File: VD073869.D

Acq: 19 Jul 2022 19:13




Tgt Ion:113 Resp: 40482


Ion Ratio Lower Upper

113 100

111 103.8 83.8 125.6

192 17.5 14.6 21.8









#50
Toluene-d8
Concen: 47.810 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VD073869.D
Acq: 19 Jul 2022 19:13




Instrument :
MSVOA_D
ClientSampleId :
SB-04-6.5-7.0


Tgt Ion: 98 Resp: 131410
Ion Ratio Lower Upper
98 100
100 65.7 52.4 78.6



#62
4-Bromofluorobenzene
Concen: 55.778 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.000 min
Lab File: VD073869.D
Acq: 19 Jul 2022 19:13

Tgt Ion: 95 Resp: 52711
Ion Ratio Lower Upper
95 100
174 75.4 0.0 152.8
176 70.9 0.0 145.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.632 min Scan# 1651
Delta R.T. -0.000 min
Lab File: VD073869.D
Acq: 19 Jul 2022 19:13

Instrument :
MSVOA_D
ClientSampleId :
SB-04-6.5-7.0

Tgt Ion:117 Resp: 106957
Ion Ratio Lower Upper
117 100
82 55.5 45.3 67.9
119 30.9 26.6 40.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. -0.000 min
Lab File: VD073869.D
Acq: 19 Jul 2022 19:13

Tgt Ion:152 Resp: 53192
Ion Ratio Lower Upper
152 100
115 61.6 31.3 93.8
150 156.9 0.0 348.6

