

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071923\
 Data File : VD076790.D
 Acq On : 19 Jul 2023 14:58
 Operator : KP/SY
 Sample : 03655-01
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MARION-COMP

Quant Time: Jul 20 04:59:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	48917	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	113971	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	113983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	49026	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	42682	87.602	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 175.200%#	
35) Dibromofluoromethane	7.804	113	43480	60.709	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 121.420%	
50) Toluene-d8	10.269	98	144503	56.400	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 112.800%	
62) 4-Bromofluorobenzene	12.575	95	42642	51.447	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 102.900%	
Target Compounds						
16) Acetone	4.022	43	2851	24.501	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071923\
Data File : VD076790.D
Acq On : 19 Jul 2023 14:58
Operator : KP/SY
Sample : 03655-01
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MARION-COMP

Quant Time: Jul 20 04:59:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

