



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD071923\
Data File : VD076782.D
Acq On : 19 Jul 2023 09:29
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Title : SW846 8260
Last Update : Mon Jul 17 08:07:52 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1858

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	15689	PASS
75	95	30	60	49.9	43765	PASS
95	95	100	100	100.0	87749	PASS
96	95	5	9	6.6	5811	PASS
173	174	0.00	2	1.1	787	PASS
174	95	50	100	85.0	74627	PASS
175	174	5	9	7.5	5634	PASS
176	174	95	101	97.5	72784	PASS
177	176	5	9	6.6	4771	PASS