

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
 Data File : VD069763.D
 Acq On : 20 Jul 2021 12:02
 Operator : VA/SY
 Sample : M3039-09
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-04-8.0-8.5

Quant Time: Jul 21 03:23:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	76915	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	138721	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	178947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	83951	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	68458	71.282	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.560%
35) Dibromofluoromethane	7.920	113	50790	54.575	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.140%
50) Toluene-d8	10.338	98	185080	51.190	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.380%
62) 4-Bromofluorobenzene	12.626	95	78038	64.717	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	129.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
Data File : VD069763.D
Acq On : 20 Jul 2021 12:02
Operator : VA/SY
Sample : M3039-09
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-04-8.0-8.5

Quant Time: Jul 21 03:23:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

