

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
 Data File : VD069773.D
 Acq On : 20 Jul 2021 16:48
 Operator : VA/SY
 Sample : M3038-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-06-5.5-6.0

Quant Time: Jul 21 03:26:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	8501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	15732	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	16916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	7274	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	6416	60.445	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.900%
35) Dibromofluoromethane	7.903	113	4168	39.491	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	78.980%
50) Toluene-d8	10.332	98	17325	42.253	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	84.500%
62) 4-Bromofluorobenzene	12.626	95	6690	48.921	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
Data File : VD069773.D
Acq On : 20 Jul 2021 16:48
Operator : VA/SY
Sample : M3038-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-06-5.5-6.0

Quant Time: Jul 21 03:26:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

