

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
 Data File : VD069774.D
 Acq On : 20 Jul 2021 17:17
 Operator : VA/SY
 Sample : M3038-03
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-07-5.5-6.0

Quant Time: Jul 21 03:27:05 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	62	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.873	114	167	50.000	ug/l	0.01
63) Chlorobenzene-d5	0.000	117	0	0.000	ug/l	-11.64
72) 1,4-Dichlorobenzene-d4	0.000	152	0	0.000	ug/l	-13.57

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0	0.000	ug/l	
Spiked Amount	50.000	Range	49 - 140	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	12.873	95	134	92.309	ug/l	0.25
Spiked Amount	50.000	Range	25 - 144	Recovery	=	184.620%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
Data File : VD069774.D
Acq On : 20 Jul 2021 17:17
Operator : VA/SY
Sample : M3038-03
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-07-5.5-6.0

Quant Time: Jul 21 03:27:05 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

