

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
 Data File : VD069776.D
 Acq On : 20 Jul 2021 18:14
 Operator : VA/SY
 Sample : M3038-07
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-11-5.5-6.0

Quant Time: Jul 21 03:27:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	43127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	74896	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	98744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	46984	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	40929	76.006	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	152.020%
35) Dibromofluoromethane	7.914	113	25215	50.183	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.360%
50) Toluene-d8	10.332	98	97801	50.101	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.200%
62) 4-Bromofluorobenzene	12.626	95	42922	65.929	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	131.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
Data File : VD069776.D
Acq On : 20 Jul 2021 18:14
Operator : VA/SY
Sample : M3038-07
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-11-5.5-6.0

Quant Time: Jul 21 03:27:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

