

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
 Data File : VD069777.D
 Acq On : 20 Jul 2021 18:42
 Operator : VA/SY
 Sample : M3039-03
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-01-9.5-10.0

Quant Time: Jul 21 03:27:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	460183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	882415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1162098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	564879	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	291297	50.696	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.400%
35) Dibromofluoromethane	7.914	113	284481	48.055	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.100%
50) Toluene-d8	10.338	98	1291306	56.146	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.300%
62) 4-Bromofluorobenzene	12.626	95	543452	70.851	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	141.700%
Target Compounds						Qvalue
16) Acetone	4.162	43	8156	19.221	ug/l	# 86
20) Methylene Chloride	4.968	84	96038	7.519	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
Data File : VD069777.D
Acq On : 20 Jul 2021 18:42
Operator : VA/SY
Sample : M3039-03
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-01-9.5-10.0

Quant Time: Jul 21 03:27:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

