


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072021\
Data File : VD069758.D
Acq On : 20 Jul 2021 08:40
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Title : SW846 8260
Last Update : Wed Jul 14 06:00:31 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.3	76485	PASS
75	95	30	60	48.9	216149	PASS
95	95	100	100	100.0	442069	PASS
96	95	5	9	6.7	29637	PASS
173	174	0.00	2	1.4	4563	PASS
174	95	50	100	71.9	317717	PASS
175	174	5	9	7.6	24157	PASS
176	174	95	101	95.3	302805	PASS
177	176	5	9	6.1	18358	PASS