

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
 Data File : VD069782.D
 Acq On : 21 Jul 2021 11:01
 Operator : VA/SY
 Sample : VD0721SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0721SBL01

Quant Time: Jul 22 03:38:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	592560	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1121115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1315763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	633938	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	324016	43.793	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.580%
35) Dibromofluoromethane	7.914	113	352572	46.876	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.760%
50) Toluene-d8	10.338	98	1499699	51.324	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.640%
62) 4-Bromofluorobenzene	12.626	95	639533	65.625	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	131.240%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.991	85	4915	1.409	ug/l	83
16) Acetone	4.168	43	5791	16.493	ug/l #	72
17) Carbon Disulfide	4.415	76	15576	1.112	ug/l #	93
39) Methylcyclohexane	9.355	83	6471	2.408	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
Data File : VD069782.D
Acq On : 21 Jul 2021 11:01
Operator : VA/SY
Sample : VD0721SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0721SBL01

Quant Time: Jul 22 03:38:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

