

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
 Data File : VD069787.D
 Acq On : 21 Jul 2021 13:23
 Operator : VA/SY
 Sample : M3094-02
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1A

Quant Time: Jul 22 03:40:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	3870	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.867	114	7147	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	10241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	3913	50.000	ug/l	-0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	3153	65.250	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.500%
35) Dibromofluoromethane	7.914	113	2549	53.162	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.320%
50) Toluene-d8	10.343	98	10204	54.779	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.560%
62) 4-Bromofluorobenzene	12.632	95	8721	140.377	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	280.760%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
Data File : VD069787.D
Acq On : 21 Jul 2021 13:23
Operator : VA/SY
Sample : M3094-02
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1A

Quant Time: Jul 22 03:40:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

