

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
 Data File : VD069792.D
 Acq On : 21 Jul 2021 15:46
 Operator : VA/SY
 Sample : M3089-14
 Misc : 4.59G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 R99997-D-BOTTOM-(3)

Quant Time: Jul 22 03:41:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	277262	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	518535	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	681979	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	346315	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	207340	59.891	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.780%
35) Dibromofluoromethane	7.914	113	183957	52.880	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.760%
50) Toluene-d8	10.338	98	703434	52.049	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.100%
62) 4-Bromofluorobenzene	12.626	95	328376	72.853	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	145.700%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
Data File : VD069792.D
Acq On : 21 Jul 2021 15:46
Operator : VA/SY
Sample : M3089-14
Misc : 4.59G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
R99997-D-BOTTOM-(3)

Quant Time: Jul 22 03:41:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

