

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
 Data File : VD069794.D
 Acq On : 21 Jul 2021 16:43
 Operator : VA/SY
 Sample : M3135-01
 Misc : 4.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-A

Quant Time: Jul 22 03:42:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 14 06:00:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	536168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1015822	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1202438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	566047	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	294227	43.949	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.900%
35) Dibromofluoromethane	7.914	113	319228	46.842	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.680%
50) Toluene-d8	10.338	98	1361486	51.423	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.840%
62) 4-Bromofluorobenzene	12.626	95	566828	64.193	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	128.380%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.962	84	170054	14.118	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
Data File : VD069794.D
Acq On : 21 Jul 2021 16:43
Operator : VA/SY
Sample : M3135-01
Misc : 4.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-A

Quant Time: Jul 22 03:42:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 14 06:00:31 2021
Response via : Initial Calibration

