



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072121\
 Data File : VD069780.D
 Acq On : 21 Jul 2021 08:47
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071321S.M
 Title : SW846 8260
 Last Update : Wed Jul 14 06:00:31 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1809

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	66581	PASS
75	95	30	60	49.1	182635	PASS
95	95	100	100	100.0	371904	PASS
96	95	5	9	6.8	25464	PASS
173	174	0.00	2	1.5	4124	PASS
174	95	50	100	73.3	272533	PASS
175	174	5	9	7.3	19760	PASS
176	174	95	101	96.1	261909	PASS
177	176	5	9	6.7	17619	PASS