


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072220\
Data File : VD065992.D
Acq On : 22 Jul 2020 09:27
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
Title : SW846 8260
Last Update : Tue Jul 21 12:34:13 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.6	29176	PASS
75	95	30	60	48.7	80805	PASS
95	95	100	100	100.0	165845	PASS
96	95	5	9	6.9	11477	PASS
173	174	0.00	2	1.0	1478	PASS
174	95	50	100	90.7	150485	PASS
175	174	5	9	7.9	11827	PASS
176	174	95	101	96.2	144736	PASS
177	176	5	9	7.4	10717	PASS