

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072321\
 Data File : VD069828.D
 Acq On : 23 Jul 2021 14:53
 Operator : VA/SY
 Sample : M3117-09
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NJ-CLEAN-9

Quant Time: Jul 24 01:34:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	13079	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	20096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	22824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	8140	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	9346	62.133	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 124.260%	
35) Dibromofluoromethane	7.914	113	6411	50.606	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.220%	
50) Toluene-d8	10.338	98	25629	52.466	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.940%	
62) 4-Bromofluorobenzene	12.626	95	8519	52.182	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 104.360%	
Target Compounds						
20) Methylene Chloride	4.961	84	3239	11.605	ug/l	# 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072321\
Data File : VD069828.D
Acq On : 23 Jul 2021 14:53
Operator : VA/SY
Sample : M3117-09
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NJ-CLEAN-9

Quant Time: Jul 24 01:34:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

