

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072321\
 Data File : VD069831.D
 Acq On : 23 Jul 2021 16:19
 Operator : VA/SY
 Sample : M3117-03RE
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NJ-CLEAN-3RE

Quant Time: Jul 24 01:35:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	73272	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	126281	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	146443	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	65586	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	61954	73.519	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.040%
35) Dibromofluoromethane	7.908	113	43461	54.594	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	109.180%
50) Toluene-d8	10.337	98	159015	51.803	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.600%
62) 4-Bromofluorobenzene	12.626	95	61014	59.475	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	118.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072321\
Data File : VD069831.D
Acq On : 23 Jul 2021 16:19
Operator : VA/SY
Sample : M3117-03RE
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NJ-CLEAN-3RE

Quant Time: Jul 24 01:35:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

