

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072321\
 Data File : VD069834.D
 Acq On : 23 Jul 2021 17:44
 Operator : VA/SY
 Sample : M3117-06
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NJ-CLEAN-6

Quant Time: Jul 24 01:35:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	14627	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	23990	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	25426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	9055	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	12981	77.165	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	154.340%
35) Dibromofluoromethane	7.908	113	7822	51.722	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.440%
50) Toluene-d8	10.338	98	29600	50.759	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.520%
62) 4-Bromofluorobenzene	12.626	95	9410	48.284	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD072321\
Data File : VD069834.D
Acq On : 23 Jul 2021 17:44
Operator : VA/SY
Sample : M3117-06
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NJ-CLEAN-6

Quant Time: Jul 24 01:35:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

