


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD072420\
 Data File : VD066037.D
 Acq On : 24 Jul 2020 10:16
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072120S.M
 Title : SW846 8260
 Last Update : Tue Jul 21 12:34:13 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	26586	PASS
75	95	30	60	51.7	75392	PASS
95	95	100	100	100.0	145920	PASS
96	95	5	9	7.3	10604	PASS
173	174	0.00	2	0.9	1267	PASS
174	95	50	100	96.1	140272	PASS
175	174	5	9	7.9	11140	PASS
176	174	95	101	95.4	133850	PASS
177	176	5	9	7.4	9947	PASS